

Allethrin Analysis in Mosquito incense

1. Analytical Condition

YL9100 HPLC

Mobile Phase : h-Hexane : IPA = 99 : 1 (1 bottle mixture)

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : UV/Vis 236nm

Column : Sil (4.6*250mm, 5um)

Injection Volume : 5ul

2. Chromatogram

