

Benzoic acid Analysis

1. Analytical Condition

YL 9100 HPLC

Column : Aegispak C18L (4.6*150mm, 5um)

Mobile Phase : Water : ACN = 55 : 45 (0.1% H₃PO₄)

Detector : UV/Vis 240nm

Flow : 1.0ml/min

Injection volume : 10μℓ

Oven Temp : 40°C

2. Chromatogram

