


## Bifenthrin Analysis



### 1. Analytical Condition

#### YL9100 HPLC

Mobile Phase : ACN : Water =80:20 (mix in one bottle)

Flow Rate : 1ml/min

Column Oven : 45°C

Detector : UV/Vis 254nm

Column : Sunfire C18 (4.6\*250mm, 5um)

Injection Volume : 10ul

### 2. Chromatogram

