

Choline Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : Water : MeOH=80:20(add 0.2% Acetic acid) => (Mix in one bottle)

Flow Rate : 1ml/min

Column Oven : 35°C

Detector : RID 35 °C

Column : Hypersil C18 (4.6*200mm, 5µm)

Injection Volume : 20µl

2. Chromatogram

