



Decursin in Angelica Extract

1. Analytical Condition

YL9100 HPLC System

Mobile Phase : 1.5 g Sodium lauryl sulfate + 2 g Sodium phosphate dibasic
+ 500 mL Acetonitrile + 500 mL Water

Flow rate : 1 mL/min

Column : Sunfire C18 (4.6 * 250 mm, 5 μ m)

Detector : UV/Vis, 230 nm

Column Temp. : 35 $^{\circ}$ C

Injection Volume: 5 μ L

2. Chromatogram

