

Hydroquinone Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : 10mM KH_2PO_4 aqueous solution : ACN = 92 : 8 (mixed in a bottle)

Flow Rate : 1ml/min

Column Oven : 40°C

Detector : UV/Vis 290nm

Column : Sunfire C18 (4.6*250mm, 5 μm)

Injection Volume : 20 μl

2. Chromatogram

