

Hydroquinone & Quinone Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : Water : ACN = 70 : 30 (included 0.1% H₃PO₄)

Flow Rate : 1ml/min

Column Oven : 40°C

Detector : UV/Vis 280nm

Column : Atlantis dC18 (4.6*250mm, 5um)

Injection Volume : 20ul

2. Chromatogram

