

Lutein Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : n-Hexane : Ethyl Acetate = 70 : 30 (mixed together in a bottle)

Flow Rate : 1.5ml/min

Column Oven : 40°C

Detector : UV/Vis 446nm

Column : Unison Silica (4.6*250mm, 3um)

Injection Volume : 10ul

2. Chromatogram

