

# Nickel Sulfamate(Nickel LSS) Analysis

## Plating additives

### 1. Analytical Condition

#### YL 9100 HPLC

Column : Capcell pak CN (4.6\*150mm, 5um)

Mobile Phase : A-Water(8ml HClO<sub>4</sub> per 1L)

B-ACN

Detector : UV/Vis 215nm

Flow : 1.0ml/min

Injection volume : 5 $\mu$ l

Oven Temp : 40°C

#### <Gradient Program>

Time	%A	%B
init	100	0
5min	70	30
6min	70	30
6.1min	100	0

### 2. Chromatogram

