

o-Cresol Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : A : Water

B : ACN

Flow Rate : 1.0 ml/min

Column Oven : 40°C

Detector : UV/Vis 254 nm

Column : C18 (4.6*250mm, 5um)

Injection Volume : 10ul

< Gradient Program >

Time (min)	A(%)	B(%)
0	70	30
20	0	100
25	0	100
30	70	30

2. Chromatogram

