

o-Cresolphthalein(OCPT) Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : 100% MeOH

Flow Rate : 0.4ml/min

Column Oven : 40°C

Detector : UV/Vis 254nm

Column : C18 (4.6*150mm, 3um)

Injection Volume : 5ul

2. Chromatogram

