


Palmitic Acid Analysis

1. Analytical Condition



YL 9100 Series QC Analyzer

Detector : UV/Vis 210nm

Flow rate : 1ml/min

Injection volume : 5 μ l

Oven Temp : 40°C

Mobile phase : ACN:Water=1:1 (0.1% H₃PO₄)

2. Chromatogram

