

# Polish Analysis in Plating Solution

## 1. Analytical Condition

### YL 9100 HPLC

Column Atlantis T3 (4.6\*150mm, 5um)

Mobile Phase : A-Water(8ml HClO<sub>4</sub> included per 1L)

B-ACN

Detector : UV/Vis 215nm

Flow : 1.0ml/min

Injection volume : 5 $\mu$ l

Oven Temp : 40°C

### <Gradient Program>

Time	%A	%B
init	100	0
10min	70	30
16min	70	30
16.1min	100	0

## 2. Chromatogram

