

Polycyclic Aromatic Hydrocarbons Analysis by Valve Switching System

1. Analytical Condition

YL9100 HPLC

Mobile Phase : n-Heptane

Flow Rate : 1ml/min

Column Oven : 35 °C

Detector : RID 35°C

Column : Silica (4.6*250mm, 5um)

Injection Volume : 10ul

2. Chromatogram

