

Puerarin in the Root of Kudzu Plant

1. Analytical Condition

YL9100 HPLC System

* Analytical Condition

Mobile Phase : Water : MeOH = 75 : 25

Flow Rate : 1mL/min

Column Oven : 35°C

Detector : UV/Vis 254 nm

Injection Volume : 10 μ L

Column : Shiseido Capcellpak MG C18 (4.6 * 250 mm, 5 μ m)

2. Chromatogram

