




# **Astaxanthin Analysis(Raw material)**



## **1. Analytical Condition**

**YL9100 HPLC**

Mobile Phase : MeOH:THF=95:5

Flow Rate : 1ml/min

Column Oven : 40°C

Detector : PDA 480nm

Column : Atactis dC18(4.6\*250mm, 5um)

Injection Volume : 10ul

## **2. Chromatogram**

