

Bis-ethyl hexyl oxyphenol methoxyphenyl triazine Analysis In Suncream (Tinosorb S[®])

1. Analytical Condition

YL 9100 Series QC Analyzer

Column : C8 (4.6 X 150mm, 5 μ m)

Detector : UV/Vis 342nm



Injection Volume : 20 μ l

Mobile Phase : ACN:MeOH=55:45

Flow Rate : 1ml/min

Column Oven : 35°C

2. Chromatogram



1. Bis-ethyl hexyl oxyphenol methoxyphenyl triazine (Tinosorb S[®])