

1,4-butanediol & Succinic acid Analysis

1. Analytical Condition

YL 6500 Series

Oven : 100°C(0.1min)->20°C/min->230°C(10min)

Column : HP-VOC (60m*0.32mm*1.8um)

Carrier gas : He, 2ml/min (Split ratio 20:1)

Injector : Capillary 250°C

Detector : FID 250°C

Injection Volume : 0.5ul

2. Chromatogram

