

파스 성분 중 dl-카파, l-멘톨 동시분석

분석조건

Column : DB-WAX 30 m x 0.25 mm x 0.25 μ m

Oven : 100°C(1min) -> 10 °C/min -> 160 °C(3min) -> 40 °C/min -> 240 °C(4min)

Carrier gas : He

flow : 2 mL/min (10:1)

Injector : Capillary, 260

Detector : FID, 260

Injection vol. : 1 μ L

분석결과

<Chromatogram>

